



William H. Brown
Christopher S. Foote
Brent L. Iverson
Eric Anslyn

<http://academic.cengage.com/chemistry/brown>

Chapter 14

Mass Spectrometry


William H. Brown • Beloit College

Mass Spectrometry (MS)

- ◆ **An analytical technique for measuring the mass-to-charge ratio (m/z) of ions in the gas phase.**
 - **Mass spectrometry is our most valuable analytical tool for determining accurate molecular masses.**
 - **Also can give information about structure.**
 - **Proteins can now be sequenced by MS.**

Mass Spectrometry (MS)

- ◆ Schematic of an electron ionization mass spectrometer (EI-MS).



© Brooks/Cole, Cengage Learning



A Mass Spectrometer

- ◆ **A mass spectrometer is designed to do three things**
 - **Convert neutral atoms or molecules into a beam of positive (or rarely negative) ions.**
 - **Separate the ions on the basis of their mass-to-charge (m/z) ratio.**
 - **Measure the relative abundance of each ion.**

A Mass Spectrometer

◆ Electron Ionization MS

- In the ionization chamber, the sample is bombarded with a beam of high-energy electrons.
- Collisions between these electrons and the sample result in loss of electrons from sample molecules and formation of positive ions.



Molecular Ion


- ◆ **Molecular ion (M):** A radical cation formed by removal of a single electron from a parent molecule in a mass spectrometer.
- ◆ For our purposes, it does not matter which electron is lost; radical cation character is delocalized throughout the molecule; therefore, we write the molecular formula of the parent molecule in brackets with
 - a plus sign to show that it is a cation.
 - a dot to show that it has an odd number of electrons.

Other MS techniques

- ◆ **What we have described is called electron ionization mass spectrometry (EI-MS).**
- ◆ **Other mass spectrometry techniques include**
 - **fast atom bombardment (FAB).**
 - **matrix-assisted laser desorption ionization (MALDI).**
 - **chemical ionization (CI).**
 - **electrospray ionization (ESI).**


Resolution

- ◆ **Resolution:** A measure of how well a mass spectrometer separates ions of different mass.
 - **Low resolution:** Refers to instruments capable of separating only ions that differ in nominal mass; that is ions that differ by at least 1 or more atomic mass units (amu).




Resolution

- ◆ **Resolution:** A measure of how well a mass spectrometer separates ions of different mass.
 - **High resolution:** Refers to instruments capable of separating ions that differ in mass by as little as 0.0001 amu. **Can Help Determine Molecular Formula**



amu = 118

exact mass = 118.0783



amu = 118

exact mass = 118.0590

Exact Mass takes into account mass of major isotopes to several decimal points

Isotopes

- **Virtually all elements common to organic compounds are mixtures of isotopes.**

Element	Atomic weight	Isotope	Mass (amu)	Relative Abundance
hydrogen	1.0079	^1H	1.00783	100
		^2H	2.01410	0.016
carbon	12.011	^{12}C	12.0000	100
		^{13}C	13.0034	1.11
nitrogen	14.007	^{14}N	14.0031	100
		^{15}N	15.0001	0.38
oxygen	15.999	^{16}O	15.9949	100
		^{18}O	17.9992	0.20
sulfur	32.066	^{32}S	31.9721	100
		^{34}S	33.9679	4.40
chlorine	35.453	^{35}Cl	34.9689	100
		^{37}Cl	36.9659	32.5
bromine	79.904	^{79}Br	78.9183	100
		^{81}Br	80.9163	98.0

Resolution

- $\text{C}_3\text{H}_6\text{O}$ and $\text{C}_3\text{H}_8\text{O}$ have nominal masses of 58 and 60, and can be distinguished by low-resolution MS.
- $\text{C}_3\text{H}_8\text{O}$ and $\text{C}_2\text{H}_4\text{O}_2$ both have nominal masses of 60.
- distinguish between them by high-resolution MS.

Molecular Formula	Nominal Mass	Precise Mass
$\text{C}_3\text{H}_8\text{O}$	60	60.05754
$\text{C}_2\text{H}_4\text{O}_2$	60	60.02112


Isotopes

- Some elements have isotopic ratios that can lead to distinctive patterns in the ms.

Element	Atomic weight	Isotope	Mass (amu)	Relative Abundance
hydrogen	1.0079	^1H	1.00783	100
		^2H	2.01410	0.016
carbon	12.011	^{12}C	12.0000	100
		^{13}C	13.0034	1.11
nitrogen	14.007	^{14}N	14.0031	100
		^{15}N	15.0001	0.38
oxygen	15.999	^{16}O	15.9949	100
		^{18}O	17.9992	0.20
sulfur	32.066	^{32}S	31.9721	100
		^{34}S	33.9679	4.40
chlorine	35.453	^{35}Cl	34.9689	100
		^{37}Cl	36.9659	32.5
bromine	79.904	^{79}Br	78.9183	100
		^{81}Br	80.9163	98.0


M+2 and M+1 Peaks

- ◆ The most common elements giving rise to significant M + 2 peaks are chlorine and bromine.
 - Chlorine in nature is 75.77% ^{35}Cl and 24.23% ^{37}Cl .
 - A ratio of M to M + 2 of approximately 3:1 indicates the presence of a single chlorine in a compound, as seen in the MS of chloroethane.



M+2 and M+1 Peaks

- Bromine in nature is 50.7% ^{79}Br and 49.3% ^{81}Br .
- A ratio of M to M + 2 of approximately 1:1 indicates the presence of a single bromine atom in a compound, as seen in the MS of 1-bromopropane.



© Brooks/Cole, Cengage Learning

M+2 and M+1 Peaks

- ◆ Sulfur is the only other element common to organic compounds that gives a significant M + 2 peak
 - $^{32}\text{S} = 95.02\%$ and $^{34}\text{S} = 4.21\%$
- ◆ Because M + 1 peaks are relatively low in intensity compared to the molecular ion and often difficult to measure with any precision, they are generally not useful for accurate determinations of molecular weight.

Molecular Ions and Interpreting a mass spectrum



- ◆ The only elements to give significant $M + 2$ peaks are Cl and Br.
 - If no large $M + 2$ peak is present, these elements are absent.
- ◆ Is the mass of the molecular ion odd or even?
- ◆ **Nitrogen Rule:** If a compound has
 - zero or an even number of nitrogen atoms, its molecular ion will have an even m/z value.
 - an odd number of nitrogen atoms, its molecular ion will have an odd m/z value.

Fragmentation of the Molecular Ion

- ◆ **To attain high efficiency of molecular ion formation and give reproducible mass spectra, it is common to use electrons with energies of approximately 70 eV [6750 kJ (1600 kcal)/mol].**
 - **This energy is sufficient not only to dislodge one or more electrons from a molecule, but also to cause extensive fragmentation.**
 - **These fragments may be unstable as well and, in turn, break apart to even smaller fragments.**


Fragmentation of M

- ◆ Fragmentation of a molecular ion, M, produces a radical and a cation.
 - Only the cation is detected by MS.



Fragmentation of M

- ◆ A great deal of the chemistry of ion fragmentation can be understood in terms of the formation and relative stabilities of carbocations in solution.
 - Where fragmentation occurs to form new cations, the mode that gives the most stable cation is favored.
 - The probability of fragmentation to form new carbocations increases in the order.





Alkanes

- ◆ Fragmentation tends to occur in the middle of unbranched chains rather than at the ends.
- ◆ The difference in energy among allylic, benzylic, 3° , 2° , 1° , and methyl cations is much greater than the difference among comparable radicals.
 - Where alternative modes of fragmentation are possible, the more stable carbocation tends to form in preference to the more stable radical.

Alkanes


- Mass spectrum of octane.




© Brooks/Cole, Cengage Learning

Alkanes

- Mass spectrum of methylcyclopentane.




© Brooks/Cole, Cengage Learning




What about 41?

Alkanes

- Mass spectrum of methylcyclopentane.



© Brooks/Cole, Cengage Learning





amu = 41

Alkenes

◆ Alkenes characteristically



- show a strong molecular ion peak.
- cleave readily to form resonance-stabilized allylic cations.



Alkynes

◆ Alkynes characteristically



- show a strong molecular ion peak.
- cleave readily to form the resonance-stabilized propargyl cation or substituted propargyl cations.



© Brooks/Cole, Cengage Learning


Cyclohexenes

- Cyclohexenes give a 1,3-diene and an alkene, a process that is the reverse of a Diels-Alder reaction (Chapter 24).




Ionization of Lone Pair-Containing Molecules

- Perhaps not surprisingly, when radicals will be taken out of lone-pair containing molecules, it is the lone pair that can often lose the electron.



Alcohols


- ◆ One of the most common fragmentation patterns of alcohols is loss of H₂O to give a peak which corresponds to M-18.
- ◆ Another common pattern is loss of an alkyl group from the carbon bearing the OH to give a resonance-stabilized oxonium ion and an alkyl radical.





Similar Rearrangements for Amines

Alcohols

- Mass spectrum of 1-butanol.





© Brooks/Cole, Cengage Learning




Amines

- ◆ The most characteristic fragmentation pattern of 1°, 2°, and 3° aliphatic amines is β -cleavage.





© Brooks/Cole, Cengage Learning

Carbonyls




- ◆ Carbonyls can also fragment. Propose a mechanism for this transformation

Carbonyls




Or alternatively, you could push arrows
other way to methyl radical

Aldehydes and Ketones




© Brooks/Cole, Cengage Learning




Why might you expect 43 be larger than 113?


Aldehydes and Ketones




© Brooks/Cole, Cengage Learning



Aldehydes and Ketones





Who's to say you can't draw it as a radical mechanism too?




Carboxylic Acids

- ◆ Characteristic fragmentation patterns are
 - α -cleavage to give the ion $[\text{CO}_2\text{H}]^+$ with m/z 45.
 - McLafferty rearrangement.



Carboxylic Acids

- Mass spectrum of butanoic acid.




© Brooks/Cole, Cengage Learning



What are the fragmentations at 60 and 45?

Carboxylic Acids

- Mass spectrum of butanoic acid.





© Brooks/Cole, Cengage Learning





Esters

◆ α -cleavage and McLafferty rearrangement



Aromatic Hydrocarbons

- Most show an intense molecular ion peak.
- Most alkylbenzenes show a fragment ion of m/z 91.



© Brooks/Cole, Cengage Learning

Mass Spectrometry

End Chapter 14